



CERTIFICATE OF ACCREDITATION

The ANSI National Accreditation Board

Hereby attests that

Indiana Standards Laboratory
2919 Shelby Street
Indianapolis, IN 46203-5236

Fulfills the requirements of

ISO/IEC 17025:2017

In the fields of

CALIBRATION and DIMENSIONAL MEASUREMENT

This certificate is valid only when accompanied by a current scope of accreditation document.
The current scope of accreditation can be verified at www.anab.org.

A handwritten signature in black ink, appearing to be 'J. Stine', is positioned above a horizontal line.

Jason Stine, Vice President

Expiry Date: 31 December 2026

Certificate Number: L2222



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

Indiana Standards Laboratory

2919 Shelby Street
 Indianapolis, IN 46203-5236
 Anthony J. Mason / Mark W. Cook
 (317) 787-6578 www.indianastandards.com

CALIBRATION AND DIMENSIONAL MEASUREMENT

Valid to: **December 31, 2026**

Certificate Number: **L2222**

Acoustics and Vibration

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|--|---------------------------|--|---|
| Sound Level Calibrator | 114 dB 125 Hz to 1 kHz | 0.23 dB | Comparison to GR 1562A Calibrator |
| | 114 dB 2 kHz | 0.34 dB | |
| | 94 dB 250 Hz, 1 kHz | 0.23 dB | Comparison to B&K 2206 SL Meter GRAS 42AG Sound Calibrator Agilent 34411A DMM |
| Sound Level Meters ^[1] Sound Level | 114 dB 125 Hz to 1 kHz | 0.25 dB | Comparison to GR 1562A Calibrator |
| | 114 dB 2 kHz | 0.33 dB | |
| | 94 dB 250 Hz, 1 kHz | 0.26 dB | Comparison to GRAS 42AG Sound Calibrator |
| Sound Level Meters ^[1] Linearity | (0 to 120) dB 1 kHz | 0.063 dB | Comparison to Ratio Transformer Agilent 3458A DMM |

Chemical Quantities

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|------------------------------------|--|--|--|
| pH Meters ^[1] | (4, 7, 10) pH | 0.02 pH | Comparison to Standard pH Buffers and Thermometer 0.1 °C |
| Conductivity Meters ^[1] | 10 μS/cm 100 μS/cm 1 000 μS/cm 10 000 μS/cm 100 000 μS/cm 1 412 μS/cm | 0.57 μS/cm 2.3 μS/cm 5.6 μS/cm 49 μS/cm 450 μS/cm 5.5 μS/cm | Comparison to Standard Solutions |
| Refractive Index Brix | (0.1 to 20) Brix | 0.2 % of reading | Comparison to Scale, Sugar Distilled Water |

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|-----------------------|---|--|--|
| Capacitance (Source) | (0.1 to 1) pF | 0.000 2 pF + 45 μF/F | Comparison to GR 1422-CD Capacitor GR 1615A Bridge |
| | (1 to 10) pF | 0.001 4 pF + 50 μF/F | |
| | (10 to 100) pF | 0.000 78 pF + 10 μF/F | Comparison to GR 1422-CL Capacitor GR 1615A Bridge |
| | (100 to 1 000) pF | 0.003 6 pF + 23 μF/F | Comparison to GR-1422-CB Capacitor GR-1615A Bridge |
| | (1 to 10) nF (10 to 100) nF (100 to 1 000) nF | 0.000 004 2 nF + 17 μF/F 0.000 14 nF + 29 μF/F 0.000 13 nF + 37 μF/F | Comparison to GR 1423A Capacitor GR 1615A Bridge |
| Capacitance (Measure) | (1 to 10) μF (10 to 100) μF | 220 μF/F 530 μF/F | Comparison to ISL Polaris Capacitance Decade GR 1615A Bridge |
| | (0.1 to 1) pF | 0.000 03 pF + 60 μF/F | Comparison to GR1615A Bridge GR 1403-K Capacitor |
| | (1 to 10) pF | 0.000 051 pF + 16 μF/F | Comparison to GR 1615A Bridge GR 1403-G Capacitor |

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|------------------------------------|----------------------|---|---|
| Capacitance (Measure) | (10 to 100) pF | 0.000 076 pF + 7.2 μ F/F | Comparison to GR 1615A Bridge GR 1404-B Capacitor |
| | (100 to 1 000) pF | 0.000 21 pF + 7.7 μ F/F | |
| | (1 to 10) nF | 0.000 001 9 nF + 11 μ F/F | |
| Capacitance (Measure) | (10 to 100) nF | 0.000 1 nF + 21 μ F/F | Comparison to GR 1615A Bridge GR 1615-P1 Bridge |
| Capacitance (Measure) | (100 to 1 000) nF | 0.000 41 nF + 27 μ F/F | Comparison to GR 1615A Bridge, GR 1409-T Capacitor |
| | (1 to 10) μ F | 0.000 008 6 μ F + 210 μ F/F | Comparison to GR 1615A Bridge |
| | (10 to 100) μ F | -0.000 014 μ F + 530 μ F/F | Comparison to GR 1689M RLC Bridge |
| AC Current (Source) ^[1] | (0.1 to 1) mA | 160 μ A/A + 0.3 nA | Comparison to GR 1440 Shunt, Holt 6A Thermal Transfer Standard, Agilent 3458A Multimeter |
| | (10 to < 50) Hz | | |
| | (0.1 to 1) mA | 88 μ A/A | |
| | (0.05 to 1) kHz | | |
| | (0.1 to 1) mA | 100 μ A/A + 0.5 nA | |
| | (> 1 to 5) kHz | | |
| | (0.1 to 1) mA | 240 μ A/A + 3.1 nA | |
| | (>5 to 10) kHz | | |
| | (>1 to 10) mA | 150 μ A/A + 1.5 nA | |
| | (10 to <50) Hz | | |
| | (>1 to 10) mA | 84 μ A/A + 1.2 nA | |
| (0.05 to 1) kHz | | | |
| (>1 to 10) mA | 83 μ A/A + 10 nA | | |
| (>1 to 5) kHz | | | |
| (>1 to 10) mA | 90 μ A/A + 74 nA | | |
| (>5 to 10) kHz | | | |
| AC Current (Source) ^[1] | (>10 to 100) mA | 150 μ A/A | Comparison to Holt CS1 Shunt, Holt 6A Thermal Transfer Standard, Agilent 3458A Multimeter |
| | (10 to <50) Hz | | |
| | (>10 to 100) mA | 90 μ A/A | |
| | (0.05 to 1) kHz | | |
| | (>10 to 100) mA | 96 μ A/A + 22 nA) | |
| | (>1 to 5) kHz | | |
| (>10 to 100) mA | 110 μ A/A + 36 n | | |
| (>5 to 10) kHz | | | |
| (>0.1 to 1) A | 150 μ A/A | | |
| (10 to <50) Hz | | | |

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|-------------------------------------|----------------------------------|---|---|
| AC Current (Source) ^[1] | (>0.1 to 1) A (0.05 to 1) kHz | 91 μ A/A | Comparison to Holt CS1 Shunt, Holt 6A Thermal Transfer Standard, Agilent 3458A Multimeter |
| | (>0.1 to 1) A (>1 to 5) kHz | 100 μ A/A – 0.74 μ A | |
| | (>0.1 to 1) A (>5 to 10) kHz | 160 μ A/A + 5 μ A | |
| | (>1 to 10) A (10 to <50) Hz | 160 μ A/A | |
| AC Current (Source) ^[1] | (>1 to 10) A (0.05 to 1) kHz | 96 μ A/A – 2.2 μ A | Comparison to Holt CS1 Shunt, Holt 6A Thermal Transfer Standard, Agilent 3458A Multimeter |
| AC Current (Source) ^[1] | (>1 to 10) A (>1 to 5) kHz | 150 μ A/A – 49 μ A | Comparison to Holt CS1 Shunt, Holt 6A Thermal Transfer Standard, Agilent 3458A Multimeter |
| AC Current (Source) ^[1] | (>10 to 20) A (10 to <50) Hz | 150 μ A/A + 110 μ A | Comparison to Holt CS1 Shunt, Holt 6A Thermal Transfer Standard, Agilent 3458A Multimeter |
| | (>10 to 20) A (0.05 to 1) kHz | 100 μ A/A | |
| | (>10 to 20) A (>1 to 5) kHz | 160 μ A/A | |
| AC Current (Measure) ^[1] | (0.1 to 1) mA (10 to <50) Hz | 160 μ A/A + 0.3 nA | Comparison to GR 1440 Shunt, Holt 6A Thermal Transfer Standard, Agilent 3458A Multimeter |
| | (0.1 to 1) mA (0.05 to 1) kHz | 88 μ A/A | |
| | (0.1 to 1) mA (>1 to 5) kHz | 100 μ A/A + 0.5 nA | |
| | (0.1 to 1) mA (>5 to 10) kHz | 240 μ A/A + 3.1 nA | |
| | (>1 to 10) mA (10 to <50) Hz | 150 μ A/A + 1.5 nA | |
| | (>1 to 10) mA (0.05 to 1) kHz | 84 μ A/A + 1.2 nA | |
| | (>1 to 10) mA (>1 to 5) kHz | 83 μ A/A + 10 nA | |
| | (>1 to 10) mA (>5 to 10) kHz | 90 μ A/A + 74 nA | |

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|-------------------------------------|------------------------------------|---|---|
| AC Current (Measure) ^[1] | (>10 to 100) mA (10 to <50) Hz | 150 μ A/A | Comparison to Holt CS1 Shunt, Holt 6A Thermal Transfer Standard, Agilent 3458A Multimeter |
| | (>10 to 100) mA (0.05 to 1) kHz | 90 μ A/A | |
| | (>10 to 100) mA (>1 to 5) kHz | 96 μ A/A + 22 nA | |
| | (>10 to 100) mA (>5 to 10) kHz | 110 μ A/A + 36 nA | |
| | (>0.1 to 1) A (10 to <50) Hz | 150 μ A/A | |
| AC Current (Measure) ^[1] | (>0.1 to 1) A (0.05 to 1) kHz | 91 μ A/A | Comparison to Holt CS1 Shunt, Holt 6A Thermal Transfer Standard, Agilent 3458A Multimeter |
| | (>0.1 to 1) A (>1 to 5) kHz | 100 μ A/A – 0.74 μ A | |
| AC Current (Measure) ^[1] | (>1 to 1) A (>5 to 10) kHz | 160 μ A/A – 5 μ A | Comparison to Holt CS1 Shunt, Holt 6A Thermal Transfer Standard, Agilent 3458A Multimeter |
| AC Current (Measure) ^[1] | (>1 to 10) A (10 to <50) Hz | 160 μ A/A | Comparison to Holt CS1 Shunt, Holt 6A Thermal Transfer Standard, Agilent 3458A Multimeter |
| | (>1 to 10) A (0.05 to 1) kHz | 96 μ A/A – 2.2 μ A | |
| | (>1 to 10) A (>1 to 5) kHz | 150 μ A/A – 49 μ A | |
| | (>1 to 10) A (>5 to 10) kHz | 260 μ A/A – 110 μ A | |
| | (>10 to 20) A (10 to <50) Hz | 150 μ A/A + 110 μ A | |
| | (>10 to 20) A (0.05 to 1) kHz | 100 μ A/A | |
| | (>10 to 20) A (>1 to 5) kHz | 160 μ A/A | |
| | (>10 to 20) A (>5 to 10) kHz | 250 μ A/A | |

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|--|---------------------------|---|---|
| DC Current (Source & Measure) ^[1,3] | (0 to 1) nA | 0.1 % reading + 160 fA | Comparison to Monitored Multifunction Calibrator, Agilent 3458A Multimeter, Standard Resistor |
| | (>1 to 10) nA | 55 μ A/A + 1.1 pA | |
| | (>10 to 100) nA | 1.5 μ A/A + 5.8 pA | |
| | >100 nA to 1 μ A | 0.4 μ A/A + 24 pA | |
| | (>1 to 10) μ A | 1.2 μ A/A + 62 pA | |
| | (>10 to 100) μ A | 1 μ A/A + 570 pA | |
| | >100 μ A to 1 mA | 1.1 μ A/A + 5.7 nA | |
| DC Current (Source & Measure) ^[1,3] | (>1 to 10) mA | 1 μ A/A + 57 nA | Comparison to Monitored Multifunction Calibrator Agilent 3458A Multimeter Standard Resistor |
| | (>10 to 100) mA | 1.1 μ A/A + 570 nA | |
| | >100 mA to 2 A | 4.1 μ A/A + 5.5 μ A | |
| | (>2 to 10) A | 22 μ A/A + 32 μ A | Comparison to Transconductance Amplifier, Agilent 3458A Multimeter, Standard Shunt |
| (>10 to 20) A | 49 μ A/A - 14 μ A | | |
| DC Current (Source & Measure) ^[1,3] | (>20 to 100) A | 67 μ A/A - 370 μ A | Comparison to Power Supply, Agilent 3458A Multimeter, Standard Shunt |
| DC Current (Measure) ^[1,3] | (>100 to 1 000) A | 0.53 mA + 0.1 % of reading | Comparison to Agilent 3458A Multimeter, Standard Shunt |
| | (>1 000 to 2 000) A | 0.13 % of reading + 32 mA | |
| DC Current (Simulated Source) ^[1] | (20 to 40) ADC | 0.04 A + 0.39 % of reading | Comparison to Transconductance Amplifier, Current Coil |
| | (40 to 200) ADC | 0.037 A + 0.48 % of reading | |
| | (200 to 1 000) ADC | 0.31 A + 0.32 % of reading | |
| Inductance (Source) | 100 μ H @ 1 kHz | 0.08 % of nominal | Comparison to General Radio 1482-B Inductor |

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|--|---------------------|---|--|
| Inductance (Source) | 1 mH @ 1 kHz | 0.02 % of nominal | Comparison to General Radio 1482-E Inductor |
| | 10 mH @ 100 Hz | 0.07 % of nominal | Comparison to General Radio 1482-H Inductor |
| | 10 mH @ 1 kHz | 0.02 % of nominal | |
| | 100 mH @ 100 Hz | 0.08 % of nominal | Comparison to General Radio 1482-L Inductor |
| | 100 mH @ 1 kHz | 0.02 % of nominal | |
| | 1 H @ 100 Hz | 0.07 % of nominal | Comparison to General Radio 1482-P Inductor |
| | 1 H @ 1 kHz | 0.02 % of nominal | |
| 10 H @ 100 Hz | 10 H @ 100 Hz | 0.07 % of nominal | Comparison to General Radio 1482-T Inductor |
| | 10 H @ 1 kHz | 0.02 % of nominal | |
| Inductance (Measure) | 100 μ H @ 1 kHz | 0.1 % of reading | Comparison to General Radio 1689 RLC Bridge |
| | 1 mH @ 1 kHz | 0.03 % of reading | |
| | 10 mH @ 100 Hz | 0.09 % of reading | |
| | 10 mH @ 1 kHz | 0.03 % of reading | |
| | 100 mH @ 100 Hz | 0.09 % of reading | |
| | 100 mH @ 1 kHz | 0.03 % of reading | |
| | 1 H @ 100 Hz | 0.08 % of reading | |
| | 1 H @ 1 kHz | 0.03 % of reading | |
| | 10 H @ 100 Hz | 0.09 % of reading | |
| | 10 H @ 1 kHz | 0.03 % of reading | |
| Magnetometers / Flux Meters | (0 to 20) G | 0.014 G + 1.2 % of reading | Comparison to Gauss Meter With Transverse Probe, Helmholtz Coil |
| | (20 to 200) G | 0.19 G + 1.2 % of reading | |
| | (200 to 2 000) G | 1.6 G + 0.82 % of reading | |
| | (2 000 to 20 000) G | 19 G + 0.73 % of reading | |
| Resistance Fixed Point (Source) ^[1] | 100 μ Ω | 4.2 μ Ω / Ω | Comparison to Otto Wolff 0.0001 Resistor, Guildline 9975 Comparator, Guildline 9923 Extender 1 Ω Standard |
| | 1 m Ω | 2.8 μ Ω / Ω | |
| | 10 m Ω | 2.2 μ Ω / Ω | |
| | 100 m Ω | 1.8 μ Ω / Ω | |

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|--|-----------------------|---|--|
| Resistance Fixed Point (Source) ^[1] | 1 Ω | 1.6 μΩ/Ω | Comparison to Guildline 9975 Comparator, 1 Ω Standard |
| | 10 Ω 100 Ω 1 kΩ | 1.2 μΩ/Ω 1.4 μΩ/Ω 1.2 μΩ/Ω | Comparison to Guildline 9975 Comparator, 100 Ω Standard |
| | 10 kΩ | 1.7 μΩ/Ω | Comparison to Guildline 9975 Comparator, 1 kΩ Standard |
| | 100 kΩ | 2.3 μΩ/Ω | Comparison to Guildline 9975 Comparator, 10 kΩ Standard |
| | 1 MΩ | 3.3 μΩ/Ω | Comparison to Guildline 9975 Comparator, 100 kΩ Standard |
| | 10 MΩ | 4.6 μΩ/Ω | Comparison to Agilent 3458A Multimeter, 1 MΩ Standard, 1 MΩ per step Decade |
| | 100 MΩ | 19 μΩ/Ω | Comparison to Agilent 3458A Multimeter, 10 MΩ Standard, 10 MΩ per step Decade |
| | 1 GΩ | 110 μΩ/Ω | Comparison to 1 GΩ Fixed Point Source, Agilent 3458A Multimeter, 100 MΩ Standard 100 MΩ per step Decade |
| | 10 GΩ | 0.085 % of reading | Comparison to Leeds Northrup 4232B Bridge 1 GΩ Standard 1 GΩ per step Decade |
| | 100 GΩ | 0.21 % of reading | Comparison to Wavetek 4800A Multifunction Calibrator, Agilent 3458A Multimeter, 100 GΩ Standard, 1 MΩ Standard |

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|---|---------------------------------|---|--|
| Resistance Ranges (Source) ^[1] | 100 $\mu\Omega$ to 1 m Ω | 70 $\mu\Omega/\Omega$ | Comparison to Guildline 9975A Comparator, Leeds & Northrup 4300 Milliohm Standard, 1 Ω Fixed Point Resistor |
| | (1 to 10) m Ω | -0.002 $\mu\Omega$ + 5.6 $\mu\Omega/\Omega$ | Comparison to Guildline 9975A Comparator, Leeds & Northrup 4300 Milliohm Standard |
| | (10 to 100) m Ω | 2.3 $\mu\Omega$ + 28 $\mu\Omega/\Omega$ | Comparison to ESI RS925D Decade Resistor, Leeds & Northrup 4222-B Resistor Agilent 3458A Multimeter |
| | 100 m Ω to 1 Ω | 1.1 $\mu\Omega$ + 14 $\mu\Omega/\Omega$ | Comparison to ESI RS925D Decade Resistor, Leeds & Northrup 4020-B Resistor Agilent 3458A Multimeter |
| | (1 to 10) Ω | 9.9 $\mu\Omega$ + 2.3 $\mu\Omega/\Omega$ | Comparison to ESI RS925A Decade Resistor, Agilent 3458A Multimeter, 10 Ω Fixed Point |
| | (10 to 100) Ω | 90 $\mu\Omega$ + 1.7 $\mu\Omega/\Omega$ | Comparison to ESI RS925A Decade Resistor, Agilent 3458A Multimeter, 100 Ω Fixed Point |
| | 100 Ω to 1 k Ω | 44 $\mu\Omega$ + 2.1 $\mu\Omega/\Omega$ | Comparison to ESI RS925A Decade Resistor, Agilent 3458A Multimeter, 1 k Ω Fixed Point |
| | (1 to 10) k Ω | 350 $\mu\Omega$ + 2.5 $\mu\Omega/\Omega$ | Comparison to ESI RS925A Decade Resistor, Agilent 3458A Multimeter, 10 k Ω Fixed Point |
| | (10 to 100) k Ω | 3 m Ω + 2.9 $\mu\Omega/\Omega$ | Comparison to ESI RS925A Decade Resistor Agilent 3458A Multimeter 100 k Ω Fixed Point |

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|---|------------------------------------|--|--|
| Resistance Ranges (Source) ^[1] | 100 kΩ to 1 MΩ | 140 mΩ + 3.9 μΩ/Ω | Comparison to ESI RS925A Decade Resistor Agilent 3458A Multimeter 1MΩ Fixed Point |
| | (1 to 10) MΩ (10 to 100) MΩ | 6.8 Ω + 7 μΩ/Ω 250 Ω + 59 μΩ/Ω | Comparison to Agilent 3458A Multimeter, PPM-R3-1111 Decade Resistor 1 MΩ Fixed Point |
| | 100 MΩ to 1 GΩ | -0.014 Ω + 290 μΩ/Ω | Comparison to PPM-R3-1111 Decade Resistor Leeds & Northrup 4232B Bridge |
| Resistance Fixed Point (Measure) ^[1] | 100 μΩ 1 m Ω 10 mΩ 100 mΩ | 4.2 μΩ/Ω 2.8 μΩ/Ω 2.2 μΩ/Ω 1.8 μΩ/Ω | Comparison to Otto Wolff 0.0001 Resistor Guildline 9975 Comparator, Guildline 9923 Extender 1 Ω Standard |
| | 1 Ω | 1.6 μΩ/Ω | Comparison to Guildline 9975 Comparator, 1 Ω Standard |
| | 10 Ω 100 Ω 1 kΩ | 1.2 μΩ/Ω 1.4 μΩ/Ω 1.2 μΩ/Ω | Comparison to Guildline 9975 Comparator, 100 Ω Standard |
| | 10 kΩ | 1.7 μΩ/Ω | Comparison to Guildline 9975 Comparator, 1 kΩ Standard |
| | 100 kΩ | 2.3 μΩ/Ω | Comparison to Guildline 9975 Comparator, 10 kΩ Standard |
| | 1 MΩ | 3.3 μΩ/Ω | Comparison to Guildline 9975 Comparator, 100 kΩ Standard |
| | 10 MΩ | 7.5 μΩ/Ω | Comparison to Agilent 3458A Multimeter, 10 MΩ Fixed Point Reference |
| | 100 MΩ | 61 μΩ/Ω | Comparison to Agilent 3458A Multimeter, 100 MΩ Fixed Point Reference Decade |

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|---|------------------------------|---|--|
| Resistance Fixed Point (Measure) ^[1] | 1 GΩ | 390 μΩ/Ω | Comparison to Agilent 3458A Multimeter, 1 GΩ Fixed Point Reference Decade |
| | 10 GΩ | 0.1 % of reading | Comparison to Certified Leeds Northrup 4232B Bridge, 10 GΩ Fixed Point Reference |
| | 100 GΩ | 0.21 % of reading | Comparison to Fluke 5730A Multifunction Calibrator, Agilent 3458A Multimeter, 100 GΩ Standard, 1 MΩ Standard |
| Resistance Ranges (Measure) ^[1] | (50 to 100) μΩ | 0.006 μΩ – 60.4 μΩ/Ω | Comparison to Guildline 9975 Comparator /9923 Extender |
| | 100 μΩ to 1 mΩ | 6.3 μΩ/Ω | |
| Resistance Ranges and Fixed Points (Measure) ^[1] | (1 to 10) mΩ | -0.002 μΩ + 5.6μΩ/Ω | Comparison to Guildline 9975A Comparator Leeds & Northrup 4300 Milliohm Standard |
| | (10 to 100) mΩ | 0.029 μΩ + 5.8 μΩ/Ω | Comparison to Guildline 9975 Comparator /9923 Extender |
| | 100 mΩ to 1 Ω (1 to 10) Ω | 1.1 μΩ + 14 μΩ/Ω 9.9μΩ + 2.3 μΩ/Ω | Comparison to Agilent 3458A Multimeter, 1 Ω Fixed Point |
| | (10 to 100) Ω | 90 μΩ + 1.7 μΩ/Ω | Comparison to ESI RS925D Decade Resistor Agilent 3458A Multimeter 100 Ω Fixed Point |
| | 100 Ω to 1 kΩ | 44 μΩ + 2.1 μΩ/Ω | Comparison to ESI RS925D Decade Resistor Agilent 3458A Multimeter 1 kΩ Fixed Point |
| | (1 to 10) kΩ | 350 μΩ + 2.5 μΩ/Ω | Comparison to ESI RS925D Decade Resistor Agilent 3458A Multimeter 10 kΩ Fixed Point |

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|---|--|--|---|
| Resistance Ranges and Fixed Points (Measure) ^[1] | (10 to 100) kΩ | 3 mΩ + 2.9 μΩ/Ω | Comparison to ESI RS925D Decade Resistor Agilent 3458A Multimeter 100 kΩ Fixed Point |
| | 100 kΩ to 1 MΩ | 140 mΩ + 3.9 μΩ/Ω | Comparison to ESI RS925D Decade Resistor Agilent 3458A Multimeter 1 MΩ Fixed Point |
| | (1 to 10) MΩ | 6.8 Ω + 7.0 μΩ/Ω | Comparison to PPM R3-1111 Decade Resistor Agilent 3458A Multimeter 10 MΩ Fixed Point |
| | (10 to 100) MΩ | 250 Ω + 59 μΩ/Ω | Comparison to PPM R3-1111 Decade Resistor Agilent 3458A Multimeter 100 MΩ Fixed Point |
| Resistance Ranges (Measure) ^[1] | 100 MΩ to 1 GΩ | -0.014 Ω + 290 μΩ/Ω | Comparison to PPM R3-1111 Decade Resistor Leeds & Northrup 4232B Bridge |
| Electrical Calibration of RTD Indicators ^[1] | (-200 to 0) °C (0 to 130) °C (130 to 600) °C (600 to 849) °C | 0.01 °C 0.02 °C 0.12 °C 0.16 °C | Comparison to Resistance Decade, RTD Tables |
| AC Voltage (Source) ^[1] | 1 mV 50 Hz to 1 kHz (>1 to 10) mV 50 Hz to 1 kHz (>10 to 100) mV 50 Hz to 1 kHz | 0.25 % of reading 120 μV/V 46 μV/V | Comparison to Ratio Transformer |
| AC Voltage (Source) ^[1] | (0.22 to 2.2) mV (10 to 20) Hz (0.22 to 2.2) mV (20 to 40) Hz (0.22 to 2.2) mV (40 to 20 000) Hz (0.22 to 2.2) mV (20 to 50) kHz (0.22 to 2.2) mV (50 to 100) kHz | 0.24 μV/mV + 4 μV 0.091 μV/mV + 4 μV 0.08 μV/mV + 4 μV 0.2 μV/mV + 4 μV 0.5 μV/mV + 5 μV | Comparison to Fluke 5730A Multifunction Calibrator |

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|------------------------------------|--|--|---|
| AC Voltage (Source) ^[1] | (0.22 to 2.2) mV (100 to 300) kHz | 1 μ V/mV + 10 μ V | Comparison to Fluke 5730A Multifunction Calibrator |
| | (0.22 to 2.2) mV (300 to 500) kHz | 1.4 μ V/mV + 20 μ V | |
| | (0.22 to 2.2) mV (500 to 1 000) kHz | 2.7 μ V/mV + 20 μ V | |
| | (2.2 to 22) mV (10 to 20) Hz | 0.24 μ V/mV + 4 μ V | |
| | (2.2 to 22) mV (20 to 40) Hz | 0.09 μ V/mV + 4 μ V | |
| | (2.2 to 22) mV (40 to 20 000) Hz | 0.08 μ V/mV + 4 μ V | |
| | (2.2 to 22) mV (20 to 50) kHz | 0.2 μ V/mV + 4 μ V | |
| | (2.2 to 22) mV (50 to 100) kHz | 0.5 μ V/mV + 5 μ V | |
| | (2.2 to 22) mV (100 to 300) kHz | 1.1 μ V/mV + 10 μ V | |
| | (2.2 to 22) mV (300 to 500) kHz | 1.4 μ V/mV + 20 μ V | |
| AC Voltage (Source) ^[1] | (2.2 to 22) mV (500 to 1 000) kHz | 2.7 μ V/mV + 20 μ V | Comparison to Fluke 5730A Multifunction Calibrator |
| | (22 to 220) mV (10 to 20) Hz | 0.24 μ V/mV + 12 μ V | |
| | (22 to 220) mV (20 to 40) Hz | 0.09 μ V/mV + 7 μ V | |
| | (22 to 220) mV (40 to 20 000) Hz | 0.057 μ V/mV + 7 μ V | |
| | (22 to 220) mV (20 to 50) kHz | 0.12 μ V/mV + 7 μ V | |
| | (22 to 220) mV (50 to 100) kHz | 0.31 μ V/mV + 17 μ V | |
| | (22 to 220) mV (100 to 300) kHz | 0.65 μ V/mV + 20 μ V | |
| | (22 to 220) mV (300 to 500) kHz | 1.4 μ V/mV + 25 μ V | |
| | (22 to 220) mV (500 to 10 00) kHz | 2.7 μ V/mV + 45 μ V | |
| | (0.22 to 2.2) V (10 to 20) Hz | 0.24 mV/V + 0.04 mV | |

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|------------------------------------|---------------------------------------|---|--|
| AC Voltage (Source) ^[1] | (0.22 to 2.2) V (20 to 40) Hz | 0.09 mV/V + 0.015 mV | Comparison to Fluke 5730A Multifunction Calibrator |
| | (0.22 to 2.2) V (40 to 20 000) Hz | 0.042 mV/V + 0.008 mV | |
| | (0.22 to 2.2) V (20 to 50) kHz | 0.067 mV/V + 0.01 mV | |
| | (0.22 to 2.2) V (50 to 100) kHz | 0.085 mV/V + 0.03 mV | |
| | (0.22 to 2.2) V (100 to 300) kHz | 0.31 mV/V + 0.14 mV | |
| | (0.22 to 2.2) V (300 to 500) kHz | 1 mV/V + 0.2 mV | |
| | (0.22 to 2.2) V (500 to 10 00) kHz | 1.7 mV/V + 0.3 mV | |
| | (2.2 to 22) V (10 to 20) Hz | 0.24 mV/V + 0.4 mV | |
| | (2.2 to 22) V (20 to 40) Hz | 0.09 mV/V + 0.15 mV | |
| | (2.2 to 22) V (40 to 20 000) Hz | 0.042 mV/V + 0.05 mV | |
| AC Voltage (Source) ^[1] | (2.2 to 22) V (20 to 50) kHz | 0.067 mV/V + 0.1 mV | Comparison to Fluke 5730A Multifunction Calibrator |
| | (2.2 to 22) V (50 to 100) kHz | 0.083 mV/V + 0.2 mV | |
| | (2.2 to 22) V (100 to 300) kHz | 0.25 mV/V + 0.6 mV | |
| | (2.2 to 22) V (300 to 500) kHz | 1 mV/V + 2 mV | |
| | (2.2 to 22) V (500 to 10 00) kHz | 1.5 mV/V + 3.2 mV | |
| | (22 to 220) V (10 to 20) Hz | 0.24 mV/V + 4 mV | |
| | (22 to 220) V (20 to 40) Hz | 0.09 mV/V + 1.5 mV | |
| | (22 to 220) V (40 to 20 000) Hz | 0.052 mV/V + 0.6 mV | |
| | (22 to 220) V (20 to 50) kHz | 0.08 mV/V + 1 mV | |
| | (22 to 220) V (50 to 100) kHz | 0.15 mV/V + 2.5 mV | |

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|------------------------------------|---|--|---|
| AC Voltage (Source) ^[1] | (0.25 V to 0.5) V 10 Hz (0.25 to 0.5) V 20 Hz (0.25 to 0.5) V 50 Hz to 50 kHz (0.25 to 0.5) V 50 kHz to 100 kHz (0.25 to 0.5) V 100 kHz to 500 kHz (0.25 to 0.5) V 500 kHz to 1 MHz (>0.5 to 1) V 10 Hz (>0.5 to 1) V 20 Hz (>0.5 to 1) V 50 Hz to 1 kHz (>0.5 to 1) V 1 kHz to 10 kHz | 150 μ V/V 70 μ V/V 65 μ V/V 70 μ V/V 225 μ V/V 750 μ V/V 135 μ V/V 65 μ V/V 55 μ V/V 45 μ V/V | Comparison to Fluke 5730A Multifunction Calibrator, Holt 6A Thermal Transfer Standard |
| AC Voltage (Source) ^[1] | (>0.5 to 1) V 10 kHz to 50 kHz (>0.5 to 1) V 50 kHz to 100 kHz (>0.5 to 1) V 100 kHz to 500 kHz (>0.5 to 1) V 500 kHz to 1 MHz (>1 to 10) V 10 Hz (>1 to 10) V 20 Hz (>1 to 10) V 50 Hz to 20 kHz (>1 to 10) V 20 kHz to 50 kHz (>1 to 10) V 50 kHz to 100 kHz (>1 to 10) V 100 kHz to 500 kHz | 40 μ V/V 50 μ V/V 150 μ V/V 625 μ V/V 125 μ V/V 50 μ V/V 30 μ V/V 40 μ V/V 50 μ V/V 150 μ V/V | Comparison to Fluke 5730A Multifunction Calibrator, Holt 6A Thermal Transfer Standard |

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment | | |
|--------------------------------------|------------------------------------|---|---|--------------|---|
| AC Voltage (Source) ^[1] | (>1 to 10) V 500 kHz to 1 MHz | 625 μ V/V | Comparison to Fluke 5730A Multifunction Calibrator, Holt 6A Thermal Transfer Standard | | |
| | (>10 to 50) V 10 Hz | 125 μ V/V | | | |
| | (>10 to 50) V 20 Hz | 50 μ V/V | | | |
| | (>10 to 50) V 50 Hz to 20 kHz | 30 μ V/V | | | |
| | (>10 to 50) V 20 Hz to 50 kHz | 40 μ V/V | | | |
| | (>10 to 50) V 50 Hz to 100 kHz | 50 μ V/V | | | |
| | (>10 to 50) V 100 Hz to 200 kHz | 150 μ V/V | | | |
| | (>50 to 100) V 10 Hz | 125 μ V/V | | | |
| | (>50 to 100) V 20 Hz | 50 μ V/V | | | |
| | (>50 to 100) V 50 Hz to 20 kHz | 30 μ V/V | | | |
| | AC Voltage (Source) ^[1] | (>50 to 100) V 20 kHz to 50 kHz | | 40 μ V/V | Comparison to Fluke 5730A Multifunction Calibrator, Holt 6A Thermal Transfer Standard |
| | | (>50 to 100) V 50 kHz to 100 kHz | | 50 μ V/V | |
| (>100 to 150) V 10 Hz | | 125 μ V/V | | | |
| (>100 to 150) V 20 Hz | | 50 μ V/V | | | |
| (>100 to 150) V 50 Hz to 1 kHz | | 30 μ V/V | | | |
| (>100 to 150) V 1 kHz to 10 kHz | | 40 μ V/V | | | |
| (>100 to 150) V 10 kHz to 20 kHz | | 50 μ V/V | | | |
| (>100 to 150) V 20 kHz to 50 kHz | | 65 μ V/V | | | |
| (>100 to 150) V 50 kHz to 100 kHz | | 100 μ V/V | | | |
| (>150 to 300) V 10 Hz | | 125 μ V/V | | | |

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|-------------------------------------|---------------------------------------|---|---|
| AC Voltage (Source) ^[1] | (>150 to 300) V 20 Hz | 50 μ V/V | Comparison to Fluke 5730A Multifunction Calibrator, Holt 6A Thermal Transfer Standard |
| | (>150 to 300) V 50 Hz to 1 kHz | 30 μ V/V | |
| | (>150 to 300) V 1 kHz to 10 kHz | 40 μ V/V | |
| | (>150 to 300) V 10 kHz to 20 kHz | 50 μ V/V | |
| | (>150 to 300) V 20 kHz to 50 kHz | 65 μ V/V | |
| | (>300 to 500) V 10 Hz | 125 μ V/V | |
| | (>300 to 500) V 20 Hz | 50 μ V/V | |
| | (>300 to 500) V 50 Hz to 1 kHz | 35 μ V/V | |
| | (>300 to 500) V 1 kHz to 10 kHz | 60 μ V/V | |
| | (>300 to 500) V 10 kHz to 20 kHz | 90 μ V/V | |
| AC Voltage (Source) ^[1] | (>300 to 500) V 20 kHz to 50 kHz | 110 μ V/V | Comparison to Fluke 5730A Multifunction Calibrator, Holt 6A Thermal Transfer Standard |
| | (>500 to 1 200) V 10 Hz | 125 μ V/V | |
| | (>500 to 1 200) V 20 Hz | 50 μ V/V | |
| | (>500 to 1 200) V 50 Hz to 1 kHz | 40 μ V/V | |
| | (>500 to 1 200) V 1 kHz to 10 kHz | 60 μ V/V | |
| | (>500 to 1 200) V 10 kHz to 20 kHz | 120 μ V/V | |
| | (>500 to 1 200) V 20 kHz to 50 kHz | 145 μ V/V | |
| | 1 mV 50 Hz to 1 kHz | 0.26 % of reading | |
| AC Voltage (Measure) ^[1] | (>1 to 10) mV 50 Hz to 1 kHz | 120 μ V/V | Comparison to Ratio Transformer |
| | (>10 to 100) mV 50 Hz to 1 kHz | 49 μ V/V | |

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|-------------------------------------|---------------------------------------|---|---|
| AC Voltage (Measure) ^[1] | (>10 to 100) mV 1 kHz to 20 kHz | 0.002 3 mV + 0.000 17 mV/mV | Comparison to Agilent 3458A Multimeter |
| | (>10 to 100) mV 20 kHz to 50 kHz | 0.002 9 mV + 0.000 34 mV/mV | |
| | (>10 to 100) mV 50 kHz to 100 kHz | 0.013 mV – 0.000 13 mV/mV | |
| AC Voltage (Measure) ^[1] | 0.25 V to 0.5 V 10 Hz | 150 μV/V | Comparison to Holt 6A Thermal Transfer Standard |
| | 0.25 V to 0.5 V 20 Hz | 70 μV/V | |
| | (0.25 to 0.5) V 50 Hz to 50 kHz | 65 μV/V | |
| | (0.25 to 0.5) V 50 kHz to 100 kHz | 70 μV/V | |
| | (0.25 to 0.5) V 100 kHz to 500 kHz | 225 μV/V | |
| | (0.25 to 0.5) V 500 kHz to 1 MHz | 750 μV/V | |
| | (>0.5 to 1) V 10 Hz | 135 μV/V | |
| AC Voltage (Measure) ^[1] | (>0.5 to 1) V 20 Hz | 65 μV/V | Comparison to Holt 6A Thermal Transfer Standard |
| | (>0.5 to 1) V 50 Hz to 1 kHz | 55 μV/V | |
| | (>0.5 to 1) V 1 kHz to 10 kHz | 45 μV/V | |
| | (>0.5 to 1) V 10 kHz to 50 kHz | 40 μV/V | |
| | (>0.5 to 1) V 50 kHz to 100 kHz | 50 μV/V | |
| | (>0.5 to 1) V 100 kHz to 500 kHz | 150 μV/V | |
| | (>0.5 to 1) V 500 kHz to 1 MHz | 625 μV/V | |
| | (>1 to 10) V 10 Hz | 125 μV/V | |
| | (>1 to 10) V 20 Hz | 50 μV/V | |
| | (>1 to 10) V 50 Hz to 20 kHz | 30 μV/V | |

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|-------------------------------------|-------------------------------------|---|--|
| AC Voltage (Measure) ^[1] | (>1 to 10) V 20 kHz to 50 kHz | 40 μ V/V | Comparison to Holt 6A Thermal Transfer Standard |
| | (>1 to 10) V 50 kHz to 100 kHz | 50 μ V/V | |
| | (>1 to 10) V 100 kHz to 500 kHz | 150 μ V/V | |
| | (>1 to 10) V 500 kHz to 1 MHz | 625 μ V/V | |
| | (>10 to 50) V 10 Hz | 125 μ V/V | |
| | (>10 to 50) V 20 Hz | 50 μ V/V | |
| | (>10 to 50) V 50 Hz to 20 kHz | 30 μ V/V | |
| | (>10 to 50) V 20 Hz to 50 kHz | 40 μ V/V | |
| | (>10 to 50) V 50 Hz to 100 kHz | 50 μ V/V | |
| | (>10 to 50) V 100 Hz to 200 kHz | 150 μ V/V | |
| | AC Voltage (Measure) ^[1] | (>50 to 100) V 10 Hz | |
| (>50 to 100) V 20 Hz | | 50 μ V/V | |
| (>50 to 100) V 50 Hz to 20 kHz | | 30 μ V/V | |
| (>50 to 100) V 20 kHz to 50 kHz | | 40 μ V/V | |
| (>50 to 100) V 50 kHz to 100 kHz | | 50 μ V/V | |
| (>100 to 150) V 10 Hz | | 125 μ V/V | |
| (>100 to 150) V 20 Hz | | 50 μ V/V | |
| (>100 to 150) V 50 Hz to 1 kHz | | 30 μ V/V | |
| (>100 to 150) V 1 kHz to 10 kHz | | 40 μ V/V | |
| (>100 to 150) V 10 kHz to 20 kHz | | 50 μ V/V | |

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|-------------------------------------|--|---|---|
| AC Voltage (Measure) ^[1] | (>100 to 150) V 20 kHz to 50 kHz (>100 to 150) V 50 kHz to 100 kHz (>150 to 300) V 10 Hz (>150 to 300) V 20 Hz (>150 to 300) V 50 Hz to 1 kHz (>150 to 300) V 1 kHz to 10 kHz (>150 to 300) V 10 kHz to 20 kHz (>150 to 300) V 20 kHz to 50 kHz (>300 to 500) V 10 Hz (>300 to 500) V 20 Hz | 65 $\mu\text{V/V}$ 100 $\mu\text{V/V}$ 125 $\mu\text{V/V}$ 50 $\mu\text{V/V}$ 30 $\mu\text{V/V}$ 40 $\mu\text{V/V}$ 50 $\mu\text{V/V}$ 65 $\mu\text{V/V}$ 125 $\mu\text{V/V}$ 50 $\mu\text{V/V}$ | Comparison to Holt 6A Thermal Transfer Standard |
| AC Voltage (Measure) ^[1] | (>300 to 500) V 50 Hz to 1 kHz (>300 to 500) V 1 kHz to 10 kHz (>300 to 500) V 10 kHz to 20 kHz (>300 to 500) V 20 kHz to 50 kHz (>500 to 1 200) V 10 Hz to 50 kHz (>500 to 1 200) V 20 Hz (>500 to 1 200) V 50 Hz to 1 kHz (>500 to 1 200) V 1 kHz to 10 kHz (>500 to 1 200) V 10 kHz to 20 kHz (>500 to 1 200) V 20 kHz to 50 kHz | 35 $\mu\text{V/V}$ 60 $\mu\text{V/V}$ 90 $\mu\text{V/V}$ 110 $\mu\text{V/V}$ 125 $\mu\text{V/V}$ 50 $\mu\text{V/V}$ 40 $\mu\text{V/V}$ 60 $\mu\text{V/V}$ 0.12 mV/V 0.15 mV/V | Comparison to Holt 6A Thermal Transfer Standard |

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|---|-------------------------|---|--|
| AC High Voltage (Source) | (>1 to 5) kV 60 Hz | -0.11 V + 5 V/kV | Comparison to AR 3605 Hypot, Ohm-Labs KV30A Divider, Agilent 34411A Multimeter |
| AC High Voltage (Measure) | (>1 to 5) kV 60 Hz | -0.11V + 5 V/kV | Comparison to Ohm-Labs KV30A Divider, Agilent 34411A Multimeter |
| AC High Voltage (Measure) | (>5 to 10) kV 60 Hz | 1.1 V + 4.7 V/kV | Comparison to Ohm-Labs KV30A Divider, Agilent 34411A Multimeter |
| | (>10 to 20) kV 60Hz | 0.021 kV + 0.004 2 kV/kV | |
| | (>20 to 60) kV 60Hz | 0.018 kV + 0.006 7 kV/kV | Comparison to Hipotronics KVM 200 Meter, Agilent 34411A Multimeter |
| DC Voltage Fixed Point (Source) | 10 mV | 19 μ V/V | Comparison to Fluke 732A DC Reference Standard 752A Reference Divider |
| | 100 mV | 2.6 μ V/V | |
| | 1 V | 1.1 μ V/V | |
| | 10 V | 1 μ V/V | |
| DC Voltage Ranges (Source) ^[1] | 100 V | 1.1 μ V/V | Comparison to Fluke 732A DC Reference Standard 752A Reference Divider |
| | 1 000 V | 1.4 μ V/V | |
| DC Voltage Ranges (Source) ^[1] | (0 to <10) μ V | 9.7 nV + 460 μ V/V | Comparison to Keithley 262 Low Thermal Divider Fluke 5730A Multifunction Calibrator Agilent 3458A Multimeter 1 V Fixed Point |
| | (10 to <100) μ V | 3.7 nV + 130 μ V/V | |
| | (100 to <1 000) μ V | 29 nV + 85 μ V/V | |
| DC Voltage Ranges (Source) ^[1] | (1 to <10) mV | 6 nV + 20 μ V/V | Comparison to Keithley 262 Low Thermal Divider Fluke 5730A Multifunction Calibrator, Agilent 3458A Multimeter, Keithley 182 Voltmeter, 10 V Fixed Point |

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|---|---|--|---|
| DC Voltage Ranges (Source) ^[1] | (10 to <100) mV (100 to <1 000) mV | 0.18 nV + 0.34 μV/V 0.14 nV + 0.9 μV/V | Comparison to Fluke 5730A Multifunction Calibrator, Fluke 752A Reference Divider, Agilent 3458A Multimeter, 1 V Fixed Point, 10 V Fixed Point |
| DC Voltage Ranges (Source) ^[1] | (1 to <10) V | 0.38 μV + 0.99 μV/V | Comparison to Fluke 5730A Multifunction Calibrator, Agilent 3458A Multimeter, 1 V Fixed Point, 10 V Fixed Point |
| DC Voltage Ranges (Source) ^[1] | (10 to <100) V (100 to 1 000) V | 1 μV + 1.1 μV/V -13 μV + 1.4 μV/V | Comparison to Fluke 5730A Multifunction Calibrator, Fluke 752A Reference Divider, Agilent 3458 Multimeter, 1 V Fixed Point, 10 V Fixed Point |
| DC High Voltage (Source) | (>1 to 5) kV | -0.037 V + 0.62 mV/V | Comparison to Extech 7021 Hipot, Ohm-Labs KV30A Divider, Agilent 34411A Multimeter |
| DC High Voltage (Source) | (>5 to 10) kV | -3.3 V + 1.3 mV/V | Comparison to AN/GSM-6B HV Source, Ohm-Labs KV30A Divider, Agilent 34411A Multimeter |
| DC High Voltage (Source) | (>10 to 30) kV | -9.4 V + 1.5 mV/V | Comparison to AN/GSM-6B HV Source Ohm-Labs KV30A Agilent 34411A Multimeter |
| DC High Voltage (Source) | (>30 to 50) kV (>50 to 70) kV | 0.034 kV + 5.5 mV/V -0.03 kV + 6.8 mV/V | Comparison to AN/GSM-6B HV Source Hipotronics KVM200 Meter, Agilent 34411A Multimeter, 1 MΩ Shunt |
| DC Voltage Fixed Point (Measure) | 100 mV 1 V 10 V 100 V 1 000 V | 2.6 μV/V 1.1 μV/V 1 μV/V 1.1 μV/V 1.4 μV/V | Comparison to Fluke 732A DC Reference Standard, 752A Reference Divider |

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|--|----------------------|---|--|
| DC Voltage Ranges (Measure) ^[1] | (1 to 10) mV | 0.18 μ V + 9.1 μ V/V | Comparison to Keithley 182 Voltmeter, 10 mV Fixed Point |
| | (10 to 100) mV | 0.18 μ V + 0.34 μ V/V | Comparison to Agilent 3458A Multimeter, 100 mV Fixed Point |
| DC Voltage Ranges (Measure) ^[1] | (100 to 1 000) mV | 0.14 μ V + 0.92 μ V/V | Comparison to Fluke 5730A Multifunction Calibrator, Agilent 3458 Multimeter, 1 V Fixed Point, 10 V Fixed Point Fluke 752A Reference Divider |
| DC Voltage Ranges (Measure) ^[1] | (1 to 10) V | 0.38 μ V + 0.99 μ V/V | Comparison to Fluke 5730A Multifunction Calibrator, Agilent 3458 Multimeter, 1 V Fixed Point, 10 V Fixed Point |
| DC Voltage Ranges (Measure) ^[1] | (10 to 100) V | 1 μ V + 1.1 μ V/V | Comparison to Fluke 5730A Multifunction Calibrator, Agilent 3458 Multimeter, 1 V Fixed Point, 10 V Fixed Point Fluke 752A Reference Divider |
| | (100 to 1 000) V | -13 μ V + 1.4 μ V/V | |
| DC High Voltage (Measure) | (>1 to 5) kV | -0.037V + 0.62 mV/V | Comparison to Extech 7021 Hipot, Ohm-Labs KV30A Divider, Agilent 34411A Multimeter |
| | (>5 to 10) kV | -3.3V + 1.3 mV/V | Comparison to AN/GSM-6B HV Source, Ohm-Labs KV30A Divider, Agilent 34411A Multimeter |
| | (>10 to 30) kV | -9.4 V + 1.5 mV/V | |
| | (>30 to 50) kV | 0.34 kV + 5.5 mV/V | Comparison to AN/GSM-6B HV Source, Hipotronics KVM200 Meter Agilent 34411A Multimeter, 1 M Ω Shunt |
| (>50 to 100) kV | -0.049 kV + 7.2 mV/V | | |

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|--|--|--|---|
| DC Ratio (Source) | (0 to 0.1) ratio (0.1 to 1) ratio | 0.28 $\mu\text{V/V}$ + (0.22 $\mu\text{V/V}$ of input X ratio) 0.6 $\mu\text{V/V}$ of input ratio | Comparison to Fluke 720A Kelvin Varley Divider |
| Thermocouple Simulation ^[1] | Type B (250 to 350) °C (350 to 445) °C (445 to 580) °C (580 to 750) °C (750 to 1 000) °C (1 000 to 1 820) °C | 1.1 °C 0.86 °C 0.68 °C 0.54 °C 0.45 °C 0.36 °C | Comparison to Ectron 1140A Thermocouple Simulator |
| | Type E (-270 to -245) °C (-245 to -195) °C (-195 to -155) °C (-155 to -90) °C (-90 to 15) °C (15 to 890) °C (890 to 1 000) °C | 1.4 °C 0.21 °C 0.12 °C 0.097 °C 0.086 °C 0.072 °C 0.086 °C | |
| Thermocouple Simulation ^[1] | Type J (-210 to -180) °C (-180 to -120) °C (-120 to -50) °C (-50 to 1 200) °C | 0.14 °C 0.12 °C 0.098 °C 0.087 °C | Comparison to Ectron 1140A Thermocouple Simulator |
| | Type K (-270 to -255) °C (-255 to -195) °C (-195 to -115) °C (-115 to -55) °C (-55 to 1 000) °C (1 000 to 1 372) °C | 2.5 °C 0.81 °C 0.14 °C 0.11 °C 0.089 °C 0.1 °C | |
| | Type N (-270 to -260) °C (-260 to -200) °C (-200 to -140) °C (-140 to -70) °C (-70 to 25) °C (25 to 160) °C (160 to 1 300) °C | 5.8 °C 1.2 °C 0.27 °C 0.18 °C 0.14 °C 0.12 °C 0.11 °C | |

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|---|--|--|---|
| | Type R (-50 to -30) °C (-30 to 45) °C (45 to 160) °C (160 to 380) °C (380 to 775) °C (775 to 1 768) °C | 0.78 °C 0.67 °C 0.52 °C 0.41 °C 0.38 °C 0.34 °C | |
| | Type S (-50 to -30) °C (-30 to 0) °C (0 to 250) °C (250 to 1 000) °C (1 000 to 1 400) °C (1 400 to 1 768) °C | 0.75 °C 0.68 °C 0.51 °C 0.44 °C 0.4 °C 0.37 °C | |
| Thermocouple Simulation ^[1] | Type T (-270 to -255) °C (-255 to -240) °C (-240 to -210) °C (-210 to -150) °C (-150 to -40) °C (-40 to 100) °C (100 to 400) °C | 2.1 °C 0.57 °C 0.35 °C 0.21 °C 0.14 °C 0.1 °C 0.089 °C | Comparison to Ectron 1140A Thermocouple Simulator |
| Oscilloscope Vertical Amplitude DC (1 MΩ) | (0 to 130) V (0 to 6.6) V | 29 μV + 0.29 mV/V 46 μV + 0.29 mV/V | Comparison to Fluke 5800A Oscilloscope Calibrator |
| Square Wave (1 MΩ) | 1 mV to 130 V pk-pk 10 Hz to 1 kHz | 230 μV + 0.59 mV/V pk-pk | Comparison to Fluke 5800A Oscilloscope Calibrator |
| | 1 mV to 130 V pk-pk (1 to 10) kHz | 53 μV + 2.9 mV/V pk-pk | |
| Square Wave (50 Ω) | 1 mV to 6 p 6 V pk-pk 10 Hz to 10 kHz | 310 μV + 2.8 mV/V pk-pk | |
| Pulse Risetime | 1 kHz to 10 MHz (200 to 350) ps | 120 ps | |
| Time Mark Source (1-2-5) | 2 ns to 20 ms 50 ms to 5 s | 1.2 μs/s -32 ns + 3.5 μs/s | |

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|---------------------------------|--|---|---|
| Time Mark Source (non-cardinal) | 2 ns to 5 s | 58 μ s/s | Comparison to Fluke 5800A Oscilloscope Calibrator |
| Leveled Sinewave (Source) | 50 kHz to 10 MHz 5 mV to 5.5 V | 35 μ V + 42 mV/V | Comparison to Fluke 5800A Oscilloscope Calibrator |
| | (10 to 30) MHz 5 mV to 5.5 V | 33 μ V + 42 mV/V | |
| | (>30 to 100) MHz 5 mV to 5.5 V | 73 μ V + 42 mV/V | |
| | (>100 to 250) MHz 5 mV to 5.5 V | 87 μ V + 50 mV/V | |
| | (>250 to 500) MHz 5 mV to 5.5 V | 100 μ V + 68 mV/V | |
| | (>500 to 600) MHz 5 mV to 5.5 V | 100 μ V + 78 mV/V | |
| | 600 MHz to 1 GHz 5 mV to 5.5 V | 18 mV/V | |
| Input Resistance (Measure) | (40 to 60) Ω (500 to 1 500) k Ω | 7.2 m Ω + 1.1 m Ω / Ω 17 Ω + 1.2 m Ω / Ω | Comparison to Fluke 5800A Oscilloscope Calibrator |
| Input Capacitance (Measure) | (5 to 50) pF | 0.61 pF + 57 mF/F | |

Electrical – RF/Microwave

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|---------------------|---------------|---|--|
| RF Power (Source) | 1 mW @ 50 MHz | 0.024 mW | Comparison to HP437B Power Meter |
| RF Power (Measure) | 1 mW @ 50 MHz | 0.019 mW | Comparison to HP 432A Power Meter Agilent 478A Thermistor Mount |

Electrical – RF/Microwave

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|---------------------------------------|---|--|---|
| RF Power (Absolute Measure) | 0.1 nW to 10 μ W (10 to 30) MHz | 0.000 003 3 μ W + 0.035 μ W/ μ W | Comparison to Agilent 437B Power Meter Agilent 8484A Power Sensor |
| | 0.1 nW to 10 μ W (30 to 50) MHz | 0.000 003 5 μ W + 0.032 μ W/ μ W | |
| | 0.1 nW to 10 μ W 50 MHz to 2 GHz | 0.000 003 5 μ W + 0.032 μ W/ μ W | |
| RF Power (Absolute Measure) | 0.1 nW to 10 μ W (2 to 12.4) GHz | 0.000 003 4 μ W + 0.032 μ W/ μ W | Comparison to Agilent 437B Power Meter Agilent 8484A Power Sensor |
| | 0.1 nW to 10 μ W (12.4 to 18) GHz | 0.000 003 3 μ W + 0.034 μ W/ μ W | |
| | 1 μ W to 100 mW 100 kHz to 1 MHz | 0.000 037 mW + 37 mW/W | |
| | 1 μ W to 100 mW (1 to 50) MHz | 0.000 04 mW + 31 mW/W | |
| | 1 μ W to 100 mW 50 MHz to 2 GHz | 0.000 04 mW + 31 mW/W | |
| 1 μ W to 100 mW (2 to 4.2) GHz | 0.000 04 mW + 32 mW/W | | |
| RF Power (Absolute Measure) | (1 to 10) μ W (4 to 10) GHz | 0.072 μ W + 0.028 μ W/ μ W | Comparison to HP 432A Power Meter HP 478A Thermistor Mount |
| | (10 to 30) μ W (4 to 10) GHz | 0.15 μ W + 0.03 μ W/ μ W | |
| RF Power (Absolute Measure) | (30 to 100) μ W (4 to 10) GHz | 0.47 μ W + 0.03 μ W/ μ W | Comparison to HP 432A Power Meter HP 478A Thermistor Mount |
| RF Power (Absolute Measure) | (100 to 300) μ W (4 to 10) GHz | 1.5 μ W + 0.03 μ W/ μ W | Comparison to HP 432A Power Meter HP 478A Thermistor Mount |
| | (0.3 to 1) mW (4 to 10) GHz | 0.004 7 mW + 30 mW/W | |
| | (1 to 3) mW (4 to 10) GHz | 0.015 mW + 30 mW/W | |
| | (3 to 10) mW (4 to 10) GHz | 0.044 mW + 30 mW/W | |
| RF Power (Relative Measure) | 0.1 nW to 10 μ W 10 MHz to 18 GHz | 0.000 005 2 μ W + 0.006 9 μ W/ μ W | Comparison to Agilent 437B Power Meter, Agilent 8484A Power Sensor |
| | 1 μ W to 100 mW 100 kHz to 4.2 GHz | 0.000 053 mW + 12 mW/W | |

Length – Dimensional Metrology

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|---|--|---|---|
| Cylindrical Pins, Plugs & Thread Wires ^[2] | (0 to 12.7) mm (0.005 to 0.5) in | 0.24 μm 9.6 μin | Comparison to PW LMU 175, Grade 0 Gage Blocks |
| Cylindrical Pins, Plugs & Thread Wires ^[2] | (12.7 to 330.2) mm (0.5 to 13.0) in | (0.33 + 0.000 11 <i>l</i>) μm (13 + 0.11 <i>L</i>) μin | Comparison to PW LMU 175, Grade 0 Gage Blocks |
| Spheres | (2.54 to 25.4) mm (0.1 to 1.0) in | (0.27 + 0.006 7 <i>l</i>) μm (11 + 6.7 <i>L</i>) μin | Comparison to PW LMU 175, Grade 0 Gage Blocks |
| Inside Micrometers ^[2] | (25.4 to 304.8) mm (1 to 12) in (304.8 to 609.6) mm (12 to 24) in (609.6 to 1 219.2) mm (24 to 48) in | (0.25+ 0.001 7 <i>l</i>) μm (9.8 + 1.6 <i>L</i>) μin (0.63 + 0.002 5 <i>l</i>) μm (25 + 2.5 <i>L</i>) μin (4.3 + 0.00067 <i>l</i>) μm (170 +0.67 <i>L</i>) μin | Comparison to PW LMU 175, PW 48” |
| End Measuring Rods ^[2] (Micrometer Standards) | (25.4 to 304.8) mm (1 to 12) in (304.8 to 609.6) mm (12 to 24) in (609.6 to 1 219.2) mm (24 to 48) in | (0.25+ 0.001 7 <i>l</i>) μm (9.8 + 1.6 <i>L</i>) μin (0.63 + 0.002 5 <i>l</i>) μm (25 + 2.5 <i>L</i>) μin (4.3 + 0.00067 <i>l</i>) μm (170 +0.67 <i>L</i>) μin | Comparison to PW LMU 175, PW 48” |
| Thickness Gages | (0.025 4 to 1.27) mm (0.001 to 0.05) in | 0.23 μm 9 μin | Comparison to PW LMU 175 |
| Steel Rules ^[2] | Up to 1 219.2 mm Up to 48 in | 0.19 mm 0.007 5 in | Comparison to Steel Rule |
| Gage Blocks ^[2] | (0.203 to 1.27) mm (0.008 to 0.05) in | (0.099 + 0.000 063 <i>l</i>) μm (3.9 + 0.063 <i>L</i>) μin | Comparison to Comparator Grade 0 Blocks |
| Gage Blocks ^[2] | (1.27 to 101.6) mm (0.05 to 4.0) in | 0.078 μm + 0.013 μm/mm 3.1 μin + 1.3 μin/in | Comparison to Comparator Grade 0 Blocks |
| | (101.6 to 500) mm (4 to 20) in | (0.97 + 0.002 6 <i>l</i>) μm (38 + 2.6 <i>L</i>) μin | Comparison to Comparator Master Blocks |
| Plain Ring Gages ^[2] | (1.016 to 3.175) mm (0.04 to 0.125) in | 0.28 μm 11 μin | Comparison to PW LMU 175 |
| | (3.175 to 25.4) mm (0.125 to 1) in | 0.25 μm 10 μin | |
| | (25.4 to 355.6) mm (1 to 14) in | (0.22 + 0.25 <i>l</i>) μm (8.7 to 1.87 <i>L</i>) μin | |
| Radius Gages ^[2] | (0.381 to 25.4) mm (0.015 6 to 1) in | 0.001 5 mm + 0.035 % of reading 0.000 059 in + 0.035 % of reading | Comparison to Starrett Vision System |

Length – Dimensional Metrology

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|--|---|--|--|
| Calipers & Linear Scales ^[1,2] | (0 to 304.8) mm (0 to 12) in (304.8 to 1 524) mm (12 to 60) in | (12 + 0.000 92 <i>l</i>) μm (480 + 0.92 <i>L</i>) μin (15 + 0.001 4 <i>l</i>) μm (590 + 1.4 <i>L</i>) μin | Comparison to Gage Blocks |
| Tape Measures ^[2] | Up to 30.48 m Up to 100 ft | (0.031 + 0.000 23 <i>l</i>) mm (0.001 2 + 0.000 23 <i>L</i>) in | Comparison to Steel Rule |
| | | (0.4 + 0.000 006 6 <i>l</i>) mm (0.016 + 0.000 006 6 <i>L</i>) in | Comparison to 50 ft Tape |
| Height Gages ^[2] | (25.4 to 1 016) mm (1 to 40) in | (0.044 + 0.002 5 <i>l</i>) μm (1.7 + 2.5 <i>L</i>) μin | Comparison to Gage Blocks, Surface Plate |
| Height Master & Riser Block ^[2] | (0 to 304.8) mm (0 to 12) in | (7.1 + 0.01 <i>l</i>) μm (280 + 13 <i>L</i>) μin | Comparison to Gage Blocks |
| Indicators, Digital, Dial & Test ^[1,2] | (0 to 101.6) mm (0 to 4) in | (2 + 0.004 3 <i>l</i>) μm (79 + 4.3 <i>L</i>) μin | Comparison to Indicator Checker |
| Micrometers, Outside, Depth, Bore Gages ^[1,2] | (2.54 to 101.6) mm (0.010 to 4) in | (1.4 + 0.002 4 <i>l</i>) μm (57 + 2.5 <i>L</i>) μin | Comparison to Gage Blocks |
| | (101.6 to 508) mm (4 to 24) in | (2.8 + 0.003 7 <i>l</i>) μm (110 + 3.7 <i>L</i>) μin | |
| Micrometers, High Accuracy ^[1] | (0 to 25.4) mm (0 to 1) in | 0.14 μm + 0.003 7 μm/mm 5.5 μin + 3.9 μin/in | Comparison to Grade 0 Gage Blocks |
| Thread Micrometer Setting Standard | (1 to 6) in (25 to 150) mm | 42 μin 1.1 μm | Comparison to PW LMU 175 |
| Thread Plugs Pitch Diameter | (0 to 4) in (0 to 101.6) mm | (79 + 1.71 <i>D</i>) μin (2 + 0.001 7 <i>D</i>) μm | Comparison to PW LMU 175 |
| Thread Plugs Major Diameter | (0 to 4) in (0 to 101.6) mm | (18 + 0.35 <i>D</i>) μin (0.46 + 0.00035 <i>D</i>) μm | Comparison to PW LMU 175 |
| Bubble Levels Level Vial Setting | (50 to 609.6) mm (1.96 to 24) in | 4.1 s Vial Setting | Comparison to Gage Blocks Surface Plate |
| Bubble Levels Vial Sensitivity | (50 to 609.6) mm (1.96 to 24) in | 3.9 s Vial Sensitivity | |
| Digital Protractors & Inclometers | (0 to 60) ° | 0.002 ° | Comparison to Gage Blocks Surface Plate Sine Bar |
| | 90 ° | 0.036 ° | Comparison to Surface Plate Cylindrical Square |

Length – Dimensional Metrology

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|--|--------------------------------|---|--|
| Measuring Microscopes Linearity ^[1] | (0 to 101.6) mm (0 to 4 in) | 5.1 μm 200 μin | Comparison to Gage Blocks |
| Angle | (0 to 90) ° | 2.6 min | Comparison to Angle Blocks |
| Profilometers & Surface Roughness Testers ^[1] | 16 Ra 119 Ra | 4.1 μin | Comparison to Roughness Standard |
| Optical Comparators Linearity ^[1] | (0 to 254) mm (0 to 10) in | 4.1 μm 160 μin | Comparison to Gage Blocks |
| Angular Scales | 90 ° | 36 s | Comparison to Angle Blocks |
| Surface Plates ^[1] Overall Flatness | Up to (6 x 6) ft | 0.36 μin + 0.85 μin/in | Comparison to Laser |
| Surface Plates ^[1] Local Area Flatness (Repeat Reading) | Up to 0.001 in | 30 μin | Comparison to Repeat-o-meter |

Mass and Mass Related

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|---|-------------------------------------|---|--|
| Air Speed Velocity (Anemometers, Pitot Tubes) | (1.5 to 30) m/s | 0.009 m/s + 1.1 % of reading | Comparison to Pitot Tube Manometer |
| Air Flow | (10 to 375) SCFM | 0.082 CFM + 0.44 % of reading | Comparison to Coriolis Flow Meter |
| Gas Flow (Mass & Volume Flow Meters) ^[1] | (30 to 400) SLPM | 0.2 SLPM + 0.27 % of reading | Comparison to Bell Prover |
| Gas Flow (Mass & Volume Flow Meters) ^[1] | (0.1 to 35) SLPM (0 to 100) SCCM | 0.002 SLPM + 0.13 % of reading 0.42 SCCM + 0.98 % of reading | Comparison to Piston Prover |
| Gas Flow (Balometers, Volume Flow Meters) | (200 to 2 000) SCFM | 1.4 SCFM + 0.9 % of reading | Comparison to Laminar Flow Element |
| Liquid Flow ^[1] | (1 to 151) L/min | 0.003 3 SLPM+ 0.12 % of reading | Comparison to Coriolis Flowmeter |
| | (0.1 to 60) L/h | 0.004 1 L/h + 0.36 % of reading | Comparison to Time, Weight, Density Correction Applied |

Mass and Mass Related

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|--|--|--|--|
| Liquid Flow ^[1] | (1 to 226) L/min | 0.003 1 SLPM + 0.14 % of reading | Comparison to Coriolis Flowmeter |
| Liquid Flow ^[1] | (39 to 1 556) L/min (0 to 400) gpm | 0.12 /min + 0.57 % of reading 0.033 gpm + 0.57 % of reading | Comparison to Mag Meter |
| Force - Compression | (9.8 to 4 452) mN (1 to 454) grf (4.44 to 4 440) N (1 to 1 000) lbf | 0.018 mN + 0.018 % of reading 0.0019 g + 0.018 % of reading 0.008 2 N + 0.016 % of reading 0.001 8 lbf + 0.016 % of reading | Comparison to Dead Weight |
| | (0.91 to 44.48) kN (204 to 10 000) lbf | 9.4 N + 0.035 % of reading 2.1 lbf + 0.035 % of reading | Comparison to 10 000 lb 1000 Series Digital Proving Ring |
| | (2 to 89) kN (460 to 20 000) lbf | 15 N + 0.013 % of reading 3 lbf + 0.001 1 % of reading | Comparison to 20 000 lb 1000 Series Digital Proving Ring |
| Force - Compression | (14.6 to 445) kN (3 284 to 100 000) lbf | 69 N + 0.0 02 % of reading 16 lbf + 0.0 02 % of reading | Comparison to 100 000 lb 1000 Series Digital Proving Ring |
| Force - Tension | (9.8 to 4 452) mN (1 to 454) grf (4.44 to 4 440) N (1 to 1 000) lbf | 0.018 mN + 0.018 % of reading 0.001 9 g + 0.018 % of reading 0.007 5 N + 0.02 % of reading 0.001 7 lbf + 0.02 % of reading | Comparison to Dead Weight |
| | (0.91 to 44.48) kN (204 to 10 000) lbf | 9.4 N + 0.035 % of reading 2.1 lbf + 0.035 % of reading | Comparison to 10 000 lb 1000 Series Digital Proving Ring |
| | (2 to 89) kN (460 to 20 000) lbf | 13 N + 0.001 1 N/N 3 lbf + 0.001 1 lbf/lbf | Comparison to 20 000 lb 1000 Series Digital Proving Ring |
| | (14.6 to 445) kN (3 284 to 100 000) lbf | 26 N + 0.1 1 % of reading 5.9 lbf + 0.1 1 % of reading | Comparison to 100 000 lb 1000 Series Digital Proving Ring |
| LEEB Hardness Tester ^[1] | (576, 655, 576) HLD | 12 HLD | Indirect Verification per ASTM 956A |
| Rockwell Hardness Testers ^[1] | HRC Low Medium High | 0.45 HRC | Indirect Verification per ASTM E18 |
| | HRBW Low Medium High | 0.63 HRBW | |

Mass and Mass Related

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|---|--|---|--|
| Rockwell Hardness Testers ^[1] | HRA Low Medium High | 0.55 HRA | Indirect Verification per ASTM E18 |
| | HREW Low Medium High | 0.44 HREW | |
| Weights | (0 to 3) g | 5.6 µg + 0.004 4 mg/g | Comparison to Mettler M3 Balance Class 1 Weights |
| | (0 to 200) g | 0.053 mg + 0.004 7 mg/g | Comparison to Sartorius ME215S Balance Class 1 Weights |
| | (0 to 5 000) g | 4 mg + 0.002 mg/g | Comparison to Voland Scale Class 1 Weights |
| | 200 g to 15 kg | 9.4 mg + 0.006 1 mg/g | Comparison to Mettler KA10-3 Comparator Class 1 Weights |
| | (1 to 30) kg | 17 mg + 0.004 8 mg/g | AND 30 Mass Comparator Class 1 Weights |
| | (10 to 50) kg | 0.92 mg + 0.007 4 mg/g | Comparison to Mettler KA50-2/P Comparator Class 1 Weights |
| Vacuum ^[1] | (0.001 to 10) torr | 0.003 7 torr | Comparison to Capacitance Manometer High Vacuum Pump Diffusion Pump |
| Pressure-Pneumatic Gage and Absolute Gage Only ^[1] | (18 to 1 000) psia (124 to 6 895) kPa | 0.002 3 % of reading | Comparison to Ruska 2465 Deadweight Tester or Transducers ^[1] |
| | (0.2 to 18.2) psia (1.37 to 124.1) kPa | 0.003 1 % of reading | |
| Pressure, Hydraulic Gage ^[1] | (1 000 to 15 000) psi (6.894 to 103.42) MPa | 0.006 7 % of reading | Comparison to Ruska 2485 Deadweight Tester or Portable Dead Weight Tester ^[1] |
| Manometers ^[1] | (0 to 20) inH2O | 0.003 inH2O | Comparison to Meriam Micromanometer |

Mass and Mass Related

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|--|---|---|---|
| Manometers ^[1] | (0 to 2) inH ₂ O | 0.001 5 inH ₂ O 0.04 mmH ₂ O | Comparison to Dwyer Microtector |
| Precision Balances (Resolution 0.1 mg) ^[1] | (0 to 205) g | 0.35 mg | Comparison to Standard Mass |
| Analytical Balances (Resolution 1 mg) ^[1] | (0 to 500) g | 0.8 mg | |
| Analytical Balances (Resolution 10 mg) ^[1] | (0 to 3 200) g | 8.1 mg | Comparison to Standard Mass |
| Bench Scales (Resolution 0.1 g) ^[1] | (0 to 32) kg | 69 mg | |
| Floor Scales (Resolution 0.2 kg) ^[1] | (0 to 907) kg | 0.11 kg | |
| Torque Analyzers | (0.1 to 2 712) N·m (0.1 to 2 000) lbf·ft | 0.04 % of reading | Comparison to Torque Arm Weights |
| Torque Wrenches ^[1,2] | (0.05 to 5.6) N·m (0.5 to 50) lbf·in (5.6 to 22.6) N·m (50 to 200) lbf·in (20.3 to 135.6) N·m (15 to 100) lbf·ft | (0.001 8 + 0.002 7T) N·m (0.016 + 0.002 7T) lbf·in (0.005 + 0.002 7T) N·m (0.044 + 0.002 7T) lbf·in (0.025 + 0.002 8T) N·m (0.018 + 0.002 8T) lbf·ft | Comparison to Torque Calibrator |
| Torque Wrenches ^[1,2] | (135.6 to 2 711.6) N·m (100 to 2 000) lbf·ft | (0.21 + 0.002 8T) N·m (0.15 + 0.002 8T) lbf·ft | Comparison to Torque Calibrator |
| Torque Watches ^[1] | (2 to 17) N·m (0.5 to 2.5) ozf·in (8 to 70) N·m (2 to 10) ozf·in | 0.1 N·m (0.014 ozf·in) 0.5 N·m (0.071 ozf·in) | Comparison to Torque Watch Calibrator |
| Torque Watches ^[1] | (42 to 303) N·m (6 to 43) ozf·in (211 to 1 518) N·m (30 to 215) ozf·in | 1.6 N·m (0.22 ozf·in) 6.8 N·m (0.96 ozf·in) | Comparison to Torque Watch Calibrator |
| Viscometers | < 10 cP (10 to 100) cP (100 to 1 000) cP (1 000 to 10 000) cP (10 000 to 100 000) cP | 0.3 % of reading 0.46 % of reading 0.5 % of reading 0.68 % of reading 0.71 % of reading | Comparison to Viscosity Standard Thermometer Water Bath |

Mass and Mass Related

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|---------------------|--------------------|---|---|
| Viscosity Cups | < 10 cSt | 0.98 % of reading | Comparison to Viscosity Standard Thermometer Water Bath Stop Watch |
| | (10 to 100) cSt | 1.7 % of reading | |
| | (100 to 1 000) cSt | 1.7 % of reading | |

Photometry and Radiometry

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|---|--|---|--|
| Illuminance responsivity (Illuminant A – CIE) White Light Meters | Up to 2 000 fc | 1.7 % of reading | Comparison to Radiometer White Light Detector |
| | (2 000 to 5 500) fc | 1.9% of reading | |
| | (5 500 to 14 400) fc | 55 fc + 2.5% | |
| | Up to 21 527 lux | 1.7 % of reading | |
| | (21 527 to 59 201) lux | 1.9% of reading | |
| | (59 201 to 155 000) lux | 592 fc + 2.5% of reading | |
| | (100 to 30 000) fL | 2.3 % of reading | |
| Spectral Irradiance UV-A (315 to 400) nm Black Light Meters | (100 to 2 000) $\mu\text{W}/\text{cm}^2$ | 5.8 % of reading | Comparison to Radiometer Black Light Detector |
| Photometric Sources: Correlated Color Temperature: Incandescent, Non-incandescent | (2 300 to 13 000) K | 33 K | Comparison to Spectrophotometer |

Thermodynamic

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|--|----------------|---|---|
| Humidity (Source) | (10 to 90) %RH | 0.59 % RH + 0.058 % of reading | Comparison to Thunder 2500 Chamber |
| Humidity (Measure and Source) ^[1] | (10 to 90) %RH | 1.4 %RH | Comparison to Vaisala HMP77B Humidity Probe, General Eastern C1 Humidity Generator |

Thermodynamic

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|--|--|--|---|
| Infrared Pyrometers | (35 to 50) °C (50 to 100) °C (100 to 350) °C (350 to 500) °C | 0.51 °C 0.67 °C 1.7 °C 2.6 °C | Comparison to Fluke 4181 Calibrator $\epsilon = 0.9$ to 1.0 $\lambda = (8$ to $14) \mu\text{m}$ |
| Temperature Uniformity ^[1] Type J Thermocouple | (-90 to 1 000) °C (-130 to 1 832) °F | 1.2 °C + 0.059 % of reading 2.2 °F + 0.059 % of reading | Comparison to Datalogger w/ External CJC, Thermocouple |
| | (-90 to 250) °C (-130 to 482) °F | 0.25 °C + 0.015 % of reading 0.45 °F + 0.015 % of reading | Comparison to Datalogger w/RTDs |
| Temperature Measure ^[1] | (-196 to -100) °C (-321 to -148) °F (-100 to 0) °C (-148 to 32) °F (0 to 660) °C (32 to 1 220) °F | 0.04 °C 0.07 °F 0.012 °C + 0.065 % of reading 0.019 °F + 0.065 % of reading 0.012 °C + 0.000 7 % of reading 0.022 °F + 0.000 7 % of reading | Comparison to Fluke 1594A Super Thermometer Rosemount 162CE SPRT |
| Temperature Measuring Equipment ^[1] | -196 ± 5 °C -321 ± 9 °F | 0.04 °C 0.07 °F | Comparison to Fluke 1594A Super Thermometer, Rosemount 162CE SPRT, LN2 Dewar |
| Temperature Measuring Equipment ^[1] | (-100 to 70) °C (-148 to 158) °F | 0.012 °C + 0.006 5 % of reading 0.019 °F + 0.006 5 % of reading | Comparison to Fluke 1594A Super Thermometer, Rosemount 162CE SPRT, Fluke 7013 Bath, Halocarbon |
| Temperature Measuring Equipment ^[1] | (60 to 300) °C (140 to 572) °F | 0.012 °C + 0.001 1 % of reading 0.022 °F + 0.001 1 % of reading | Comparison to Fluke 1594A Super Thermometer, Rosemount 162CE SPRT, Fluke 7013 Bath, Silicone Oil |
| | (150 to 400) °C (302 to 770) °F | 0.012 °C + 0.001 1 % of reading 0.022 °F + 0.001 1 % of reading | Comparison to Fluke 1594A Super Thermometer Rosemount 162CE SPRT Fluke 6045 Salt Bath |
| ITS 90 – Fixed Point | 660 °C 1 220 °F | 0.54 °C 0.97 °F | Comparison to Aluminum Freeze Point Thermocouple Indicator |
| Surface Temperature Measurement ^[1] | (0 to 250) °C (32 to 482) °F | 1.3 °C 2.3 °F | Comparison to Fluke 741 Process Calibrator Type K Surface Probe |
| Precision Thermometry | -196 °C | 10 mK | Comparison to SPRT |

Thermodynamic

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|-----------------------|--------------|---|--|
| Precision Thermometry | -38.834 4 °C | 1.6 mK | Comparison to Mercury Triple Point Cell |
| Precision Thermometry | 0.010 °C | 0.97 mK | Comparison to Water Triple Point Cell |
| Precision Thermometry | 29.764 6 °C | 1.3 mK | Comparison to Gallium Melting Point Cell |
| Precision Thermometry | 156.598 5 °C | 2.1 mK | Comparison to Indium Freezing Point Cell |
| Precision Thermometry | 231.928 °C | 3.8 mK | Comparison to Tin Freezing Point Cell |
| Precision Thermometry | 419.527 °C | 7.0 mK | Comparison to SPRT |
| Precision Thermometry | 660.323 °C | 11 mK | Comparison to SPRT |

Time and Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|--------------------------------------|---|---|---|
| Time Interval | (0.1 to 5) ns | 1 % reading + 15 ps | Comparison to Tektronix 2465 Oscilloscope |
| | (5 to 50) ns | 0.5 % reading + 150 ps | |
| | 3.3 ns to 10 ¹⁰ s | 2.5 x 10 ⁻⁹ Hz / Hz + 500 ps | Comparison to Fluke PM6681R Counter |
| Frequency (Measure) | 10 Hz to 200 MHz | 2.5 x 10 ⁻⁹ Hz/Hz | Comparison to Fluke PM6681R Counter |
| | (0.2 to 2) GHz | 2.6 x 10 ⁻⁹ Hz/Hz | |
| | (2 to 26) GHz | 6.2 x 10 ⁻⁹ Hz/Hz | |
| Frequency (Source) | 10 Hz to 200 MHz | 2.5 x 10 ⁻⁹ Hz/Hz | Comparison to Signal Generator monitored with Fluke PM6681R Counter |
| | (0.2 to 2) GHz | 2.6 x 10 ⁻⁹ Hz/Hz | |
| Stop Watches ^[1] | Up to 24 hr | 0.058 s/day | Comparison to Helmut Klein 4500 Timometer |
| | (1 to 3 600) s | 0.12 s | Comparison to Timer Counter |
| Tachometers (Contact) ^[1] | (5.76 to 4 189) rad/s (55 to 40 000) rpm | 0.21 rad/s + 0.007 % of reading 0.2 rpm + 0.007 % of reading | Comparison to rpm Standard |

Time and Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|---|--|--|---|
| Tachometers (Contact) ^[1] | (100 to 1 000) ft/min | 0.026 % of reading | Comparison to rpm Standard Standard Wheel |
| Tachometers (Non-contact) Strobe & Photo ^[1] | (0.62 to 10 472) rad/s (6 to 100 000) rpm | 0.000 4 rad/s + 0.000 6 % of reading 0.003 8 rpm + 0.000 6 % of reading | Comparison to Function Generator |
| Rpm (Measure) | (6 to 100 000) rpm | 1.2 rpm | Comparison to Optical Tachometer |



DIMENSIONAL MEASUREMENT

1 Dimensional

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|---------------------|--|---|--|
| Linear | Up to 304.8 mm Up to 12 in | 0.003 6 mm + 0.000 37 % of reading 0.000 14 in + 0.000 37 % of reading | Comparison to Starrett Vision System |
| Linear Measurement | (25.4 to 50 800) mm (1 to 2 000) in | 0.45 μ m + 0.57 μ m/mm 18 μ in + 0.57 μ in/in | Comparison to Laser Measuring Machine |

2 Dimensional

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|---------------------|--------------------------------|---|--|
| Angular | (0 to 360) ° | 0.001 3 ° + 0.000 77 % of reading | Comparison to Starrett Vision System |
| Linear (2D) | (254 to 127) mm (10 x 5) in | 0.003 6 mm + 0.000 37 % of reading 0.000 14 in + 0.000 37 % of reading | Comparison to Starrett Vision System |

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ($k=2$), corresponding to a confidence level of approximately 95%.

Notes:

1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
2. D = diameter in mm, DL = diagonal length in inches, fL = foot lambert, inD = diagonal inches, l = length in mm, L = length in inches, mil = 1/1000 of an inch or 0.001 inch, T = applied torque.
3. Test currents up to 1000 A are generated using 50 turn coil with no loss of accuracy.
4. This scope is formatted as part of a single document including Certificate of Accreditation No. L2222
5. Chance Indiana Standards Laboratory dba Indiana Standards Laboratory.



Jason Stine, Vice President